

**Contact Person:**

Name: Pierrot, Denis  
Organization: NOAA/Atlantic Oceanographic & Meteorological Laboratory  
Address: 4301 Rickenbacker Causeway, Miami Fl, 33149  
Phone: 305-361-4441  
Email: Denis.Pierrot@noaa.gov

**Investigator(s):**

Name: Wanninkhof, Rik  
Organization: NOAA/Atlantic Oceanographic & Meteorological Laboratory  
Address: 4301 Rickenbacker Causeway, Miami Fl, 33149  
Phone: 305-361-4379  
Email: Rik.Wanninkhof@noaa.gov

Name: Pierrot, Denis  
Organization: NOAA/Atlantic Oceanographic & Meteorological Laboratory  
Address: 4301 Rickenbacker Causeway, Miami Fl, 33149  
Phone: 305-361-4441  
Email: Denis.Pierrot@noaa.gov

**Dataset Information:**

Funding\_Info: NOAA Climate Program Office; NOAA Ocean Acidification Program  
Initial\_Submission: 20160130  
Revised\_Submission: 20160130

**Cruise Information:**

Experiment Name: GU1103\_Leg2  
Experiment Type: Research Cruise  
Platform Type: Ship  
Co2 Instrument Type: Equilibrator-IR or CRDS or GC

Cruise ID: 33GG20110712  
Cruise Info: AOML\_SOOP\_CO2  
Geographical Region:

Westernmost Longitude: -88.6  
Easternmost Longitude: -74.1  
Northernmost Latitude: 37.4  
Southernmost Latitude: 24.2

Cruise Dates (YYYYMMDD)

Start\_Date: 20110713  
End\_Date: 20110802

Ports of Call:

Pascagoula, MS  
Norfolk, VA

Vessel Name: R/V Gordon Gunter  
Vessel ID: 33GG

Vessel Owner: NOAA

**Variables Information:**

Variable Name: xCO2\_EQU\_ppm

Description of Variable: Mole fraction of CO2 in the equilibrator headspace (dry) at equilibrator temperature (ppm)

Unit of Variable: ppm

Variable Name: xCO2\_ATM\_ppm

Description of Variable: Mole fraction of CO2 measured in dry outside air (ppm)

Unit of Variable: ppm

Variable Name: xCO2\_ATM\_interpolated\_ppm

Description of Variable: Mole fraction of CO2 in outside air associated with each water analysis. These values are interpolated between the bracketing averaged good xCO2\_ATM analyses (ppm)

Unit of Variable: ppm

Variable Name: PRES\_EQU\_hPa

Description of Variable: Barometric pressure in the equilibrator headspace (hPa)

Unit of Variable: hPa

Variable Name: PRES\_ATM@SSP\_hPa

Description of Variable: Barometric pressure measured outside, corrected to sea level (hPa)

Unit of Variable: hPa

Variable Name: TEMP\_EQU\_C

Description of Variable: Water temperature in equilibrator (°C)

Unit of Variable: Degree C

Variable Name: SST\_C

Description of Variable: Sea surface temperature (°C)

Unit of Variable: Degree C

Variable Name: SAL\_permil

Description of Variable: Sea surface salinity on Practical Salinity Scale (o/oo)

Unit of Variable: ppt

Variable Name: fCO2\_SW@SST\_uatm

Description of Variable: Fugacity of CO2 in sea water at SST and 100% humidity ( $\mu\text{atm}$ )

Unit of Variable:  $\mu\text{atm}$

Variable Name: fCO2\_ATM\_interpolated\_uatm

Description of Variable: Fugacity of CO2 in air corresponding to the interpolated xCO2 at SST and 100% humidity ( $\mu\text{atm}$ )

Unit of Variable:  $\mu\text{atm}$

Variable Name: dfCO2\_uatm

Description of Variable: Sea water fCO2 minus interpolated air fCO2 ( $\mu\text{atm}$ )

Unit of Variable:  $\mu\text{atm}$

Variable Name: WOCE\_QC\_FLAG

Description of Variable: Quality control flag for fCO2 values (2=good, 3=questionable)

Unit of Variable: None

Variable Name: QC\_SUBFLAG

Description of Variable: Quality control subflag for fCO<sub>2</sub> values, provides explanation when QC flag=3

Unit of Variable: None

### Method Description:

#### Equilibrator Design:

Depth of Seawater Intake: 5 meters

Location of Seawater Intake: Bow

Equilibrator Type: Spray head above dynamic pool, no thermal jacket

Equilibrator Volume: 0.95 L (0.4 L water, 0.55 L headspace)

Water Flow Rate: 1.5 - 2.0 L/min

Headspace Gas Flow Rate: 70 - 150 ml/min

Vented: Yes

Drying Method for CO<sub>2</sub> in Water:

Gas stream passes through a thermoelectric condenser (~5 °C) and then through a Perma Pure (Nafion) dryer before reaching the analyzer (90% dry).

Additional Information: Primary equilibrator is vented through a secondary equilibrator.

#### CO<sub>2</sub> in Marine Air:

Measurement: Yes, 5 readings in a group every 3 hours

Location and Height: Bow mast, ~15 meters above sea surface

Drying Method:

Gas stream passes through a thermoelectric condenser (~5 °C) and then through a Perma Pure (Nafion) dryer before reaching the analyzer (90% dry).

#### CO<sub>2</sub> Sensor:

Measurement Method: IR

Manufacturer: LI-COR

Model: 7000

Frequency: Every 140 seconds, except during calibration

Resolution Water: ± 0.01 µatm in fCO<sub>2</sub>\_SW

Uncertainty Water: ± 2 µatm in fCO<sub>2</sub>\_SW

Resolution Air: ± 0.01 µatm in fCO<sub>2</sub>\_ATM

Uncertainty Air: ± 0.5 µatm in fCO<sub>2</sub>\_ATM

Manufacturer of Calibration Gas:

Std 2: JA02280, 248.73 ppm, owned by AOML, used every ~2.5 hours. Std 3: JA02292, 372.88 ppm, owned by AOML, used every ~2.5 hours. Std 4: JA02689, 520.79 ppm, owned by AOML, used every ~2.5 hours.

Number of Non Zero Gas Standards: 3

CO<sub>2</sub> Sensor Calibration:

The analyzer is calibrated every 4.5 hours with field standards that in turn were calibrated with primary standards that are directly traceable to the WMO scale. The zero gas is ultra-high purity air.

Other Comments:

Instrument is located in an air-conditioned laboratory. Ultra-High Purity air (0.0 ppm CO<sub>2</sub>) and the high standard gas are used to zero and span the LI-COR analyzer.

Method References:

Pierrot, D., C. Neil, K. Sullivan, R. Castle, R. Wanninkhof, H. Lueger, T. Johannessen, A. Olsen, R. A. Feely, and C. E. Cosca (2009), Recommendations for autonomous underway pCO<sub>2</sub> measuring systems and data reduction routines, Deep-Sea Res II, 56, 512-522.

Details Co2 Sensing:  
details of CO2 sensing (not required)  
Measured Co2 Params:  
xco2(dry)

Sea Surface Temperature:  
Location: hull mounted, ~3 m below sea surface  
Manufacturer: Furuno  
Model: T2000  
Accuracy Degrees Celsius: 0.2  
Precision Degrees Celsius: 0.1  
Calibration: Factory calibration  
Comments: Manufacturer's Resolution is taken as Precision; Maintained by ship.

Equilibrator Temperature:  
Location: Inserted into equilibrator ~5 cm below water level  
Manufacturer: Hart  
Model: 1521  
Accuracy Degrees Celsius: 0.025  
Precision Degrees Celsius: 0.001  
Calibration: Factory calibration  
Comments: Resolution is taken as Precision.

Equilibrator Pressure:  
Location: Attached to equilibrator headspace. Combined with Licor Pressure  
Manufacturer: Licor  
Model: None  
Accuracy hPa: 1.2  
Precision hPa: 0.02  
Calibration: Factory calibration  
Comments:  
Differential pressure reading from Setra-239 attached to the equilibrator headspace was added to the pressure reading from the LICOR analyzer to yield equilibrator pressure. Manufacturer's Resolution is taken as Precision.

Atmospheric Pressure:  
Location: Next to the bridge, ~15 m above the sea surface water  
Manufacturer: RMYoung  
Model: 61201  
Accuracy:  $\pm 0.5$  hPa  
Precision: 0.01 hPa  
Calibration: Factory calibration  
Normalized: yes  
Comments: Manufacturer's Resolution is taken as Precision; Maintained by ship.

Sea Surface Salinity:  
Location: In Chem lab, next to CO2 system  
Manufacturer: Seabird  
Model: SBE 21  
Accuracy:  $\pm 0.05$  o/oo  
Precision: 0.002 o/oo  
Calibration: Factory calibration  
Comments: Manufacturer's Resolution is taken as Precision; Maintained by ship.

**Additional Information:**

LI-7000 Firmware 2.0 bug correction was applied on xCO2 values. Zero standard was out. Only 3 standards used. All original data can be found at: [http://www.aoml.noaa.gov/ocd/ocdweb/gunter/gunter\\_introduction.html](http://www.aoml.noaa.gov/ocd/ocdweb/gunter/gunter_introduction.html)

**Preliminary Quality Control:**

NA

**Form Type:**

underway